
# **MUSSD**

### Metrological Tests Sample Acceptance/ Rejection Form: Measuring Cans

Date of Issue:	Doc No:
2023-01-01	LB/PA/VL/F2.3
Rev No: 1	Page 8 of 8

### **Annex**

## **Specifications of Measuring Cans**



↑ B B

Figure 2: Dimensions of cans



Figure 3 : can




Figure 4: Markings of cans

- Figure 1: Dimensions of cans
- 1. Over flow hole  $(5 \text{mm}, \boxtimes)$
- 2. Riveted, welded, soldered or brazed
- 3. Name and Denomination plate
- 4. Seal

#### **Key (in millimeters)**

- ) (	,										
DENOMINATION	Α	В	С	D	Е	F	G	Н	J	K	M
							Min.				
20 Liters	97	388	388	208	194	390	1.00	35	86	29	30
10 Liters	77	308	307	174	154	309	1.00	30	75	26	25
5 Liters	61	244	245	147	122	247	0.80	25	65.5	24	20
2 Liters	45	180	180	118	90	182	0.80	20	56	22	16
1 Liter	36	143	143	95.5	72	145	0.63	20	45	18	16
500 milliliters	28	114	113	74	56	115	0.63	15	35	14	12
200 milliliters	21	84	84	53	42	86	0.63	10	24.5	10	8
100 milliliters	17	66	67	41	34	69	0.63	10	18.5	7	8

<sup>\*</sup>Tolerance on dimensions:  $\pm 10$  % or  $\pm 5$  mm (whichever is higher) except in case of 10 Liter and 20 Liter measures for which the tolerance should be 5% or  $\pm 5$  mm (whichever is higher)

Prepared by: TM/DTM Reviewed by: QM/DQM Approved by: Director

<sup>\*</sup>Tolerance on angles: ± 5°